



## EuCC presents recommendations for short, angulated and reduced-diameter implants

# A reliable treatment option

In January and February, the 18<sup>th</sup> European Consensus Conference (EuCC) under the auspices of BDIZ EDI updated its 2016 Guideline on short, angulated and reduced-diameter implants. The 2023 Guideline provides recommendations for practitioners and reflects data from controlled clinical trials while also incorporating data from routine clinical practice.

Prof. Jörg Neugebauer, BDIZ EDI Secretary General and host of the EuCC, explains why this revised version has been prepared: "Discussions on this topic do not take place in a closed forum. We are not aiming for a purely academic environment. Rather, our recommendations should provide practical guidance for practising dentists, while also incorporating expertise from across Europe."

### Short implants

This second update has left the definition of "short" unchanged. As before, they have a designed intrabony length of  $\leq 8$  mm with a diameter of  $\geq 3.75$  mm. They are used, among other things, to avoid bone grafting in the posterior jaw segments of partially edentulous patients, but also to support removable overdentures and as single or multiple tooth replacements in the anterior jaw. The EuCC

has found that there is no longer any difference in success rates compared with standard implants with augmentation procedures. A new indication is that for immediate loading. There are now studies that support the use of short implants with special treatment concepts in immediate-loading situations.

### Angulated implants

There have also been new developments regarding angulated implants, which are becoming routine in splinted reconstructions of edentulous jaws. The EuCC agreed that they increase primary stability for immediate loading procedures

## Bibliographical note

### Guideline 2023

"Update on short, angulated and reduced-diameter implants"

Prepared by the 18<sup>th</sup> European Consensus Conference under the auspices of BDIZ EDI in January and February 2023

11 A4 pages + cover, with extensive references

Come to BDIZ EDI stand at IDS: Hall 11.2, aisle O, stand 69 and get your free copy. Additional copies can be ordered from the BDIZ EDI online store at [www.bdizedi.org](http://www.bdizedi.org) (€3.50 incl. VAT, plus S + H)



are significantly lower, but patients will still benefit in terms of oral health-related quality of life. Mini implants also show favourable results when used to increase the number of restorative abutments for removable partial dentures. According to the EuCC, short mini implants should still be avoided.

less risky therapeutic option in terms of specific treatment parameters, compared with the risks associated with the use of standard-dimension implants in combination with augmentation procedures.

AWU

## Recommendation

The use of short, angulated or reduced-diameter implants in sites with reduced bone volume can be a reliable, faster and

with longer implants, avoiding bone grafting. These treatment concepts require 4 implants in the mandible and 4 to 6 implants in the maxilla. However, current observations have also revealed limitations. "Despite the positive clinical results, the scientific debate on the clinical relevance of the development of marginal bone levels around angulated implants is still ongoing", as Neugebauer summarised the consensus finding.

## Reduced-diameter implants

The EuCC distinguishes between two general settings. Reduced-diameter implants—those with intraosseous diameters of < 3.5 mm—are indicated for use in jaws with reduced widths. EuCC refers to implants with diameters of < 2.7 mm as mini implants. There is no change from to the previous 2016 Guideline. New meta-analyses support the statements made at that time.

Reduced-diameter implants have high survival rates (> 90%) with careful patient selection, bone density assessment, clinical approach, and user experience. They can also be used in the posterior region with high success rates.

There are differences in the success rates of mini implants in the maxilla and mandible. While mini implants in the mandible that are restored with an overdenture have excellent short- to medium-term survival rates, survival rates in the maxilla



### Guideline 2023

#### 2<sup>nd</sup> Update on short, angulated and reduced-diameter implants

#### 18th European Consensus Conference (EuCC) 2023

6 February 2023

**Authors:** Jörg Neugebauer, PhD, DMD  
Joachim E. Zöller, PhD, MD, DMD  
Department of Craniomaxillofacial and Plastic Surgery  
and Interdisciplinary Department for Oral Surgery and Implantology  
Centre for Dentistry and Oral and Maxillofacial Surgery,  
University of Cologne, Germany

**Host:** Professor J. Neugebauer (Germany)  
**Protocol:** Professor J. Neugebauer (Germany)

**Participants:** Professor B. Al-Nawas (Germany)  
C. Berger (Germany)  
Dr E. O'Connell (United Kingdom)  
Professor A. Felino (Portugal)  
Dr I. Frank (Germany)  
Professor F. Heinemann (Germany)  
Dr F. Kasapi (Macedonia)  
Dr K. Krasny (Poland)  
Professor V. Konstantinović (Serbia)  
Dr S. Liepe (Germany)  
Dr A. Manolakis (Greece)  
Professor J. Neugebauer (Germany)  
Dr W. Neumann (Germany)  
Professor H. Özyuvacı (Turkey)  
Dr B. Singh (Nepal)  
W. Tomkiewicz (Poland)  
Dr J.W. Vaartjes (The Netherlands)  
Professor J.E. Zöller (Germany)

### Content

1 Methods	Page 2
2 Problem	Page 3
3 Use of short implants	Page 3
4 Use of angulated implants	Page 4
5 Use of reduced-diameter implants	Page 5
6 Recommendations	Page 5
7 References	Page 6

Bundesverband der  
implantologisch  
tätigen Zahnärzte  
in Europa

European  
Association of  
Dental  
Implantologists

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org



## 1 Methods

### 1.1 Objective

The purpose of this guideline is to provide clinicians involved in implant dentistry with recommendations to enable them to correctly assess the potential indications (and any limitations) for short, angulated or reduced-diameter implants. This guideline reflects data from controlled clinical trials and takes into account clinical data from routine treatment in the private-practice setting.

### 1.2 Introduction

This consensus paper is concerned only with titanium implants, typically placed according to the indications recommended by the European Consensus Conference Implantology (EuCC, Germany, 6 February 2023).

All consensus recommendations in this paper should be considered as guidelines only. The specific situation of the patient is always an important consideration and may justify a deviation from the recommendations of this consensus paper.

### 1.3 Background

Avoiding bone augmentation with reduced-dimension implants and making optimal use of the available bone volume are often recommended as minimally invasive treatment options [50]. In addition to the number of implants, dimension and insertion type must be considered to ensure an acceptable treatment outcome.

### 1.4 Literature search

The Cochrane Library, EMBASE, DIMDI and Medline literature databases were used to conduct a systematic search for recently published data on the use of short, angled and reduced-diameter implants. Selective search criteria were used, including terms such as short implants, angulated implants, angled implants, tilted implants and implant failure, narrow and reduced diameter. The publications identified by the search were screened by reading their abstracts, and those irrelevant to the topic were identified and excluded. Those articles identified as potentially relevant were obtained in full text. Several meta-analysis reviews and randomized controlled trials (RCTs) and other prospective or retrospective systematic clinical trials were available on the topic.

### 1.5 Development of this guideline/consensus paper

A preliminary version, on which the EuCC based its deliberations, was prepared and reviewed by Professor J. Neugebauer of the Interdisciplinary Policlinic for Oral Surgery and Implantology and the Department of Oral and Maxillofacial Plastic Surgery at the University of Cologne, Germany. The preliminary report was then reviewed and discussed by the members of the committee in the following five steps:

- Review of the preliminary draft
- Collecting alternative suggestions
- Voting on recommendations and levels of recommendation
- Discussion of non-consensual issues
- Final vote

BDIZ EDI  
 Lipowskystr. 12  
 D-81373 Munich  
 GERMANY

Fon: +49-89-720 69 888  
 Fax: +49-89-720 69 889  
 office@bdizedi.org  
 www.bdizedi.org



## 2 Problem

The use of standard implants in patients with alveolar ridge atrophy or extensive pneumatization of the maxillary sinus cavity often requires the use of hard-tissue augmentation procedures [18, 19]. These procedures are well established, and widely used with success. However, depending on the operator's level of training and the patient-specific risk factors complications may occur, or the postoperative quality of life may be compromised [2, 10, 17-19, 34].

## 3 Use of short implants

### 3.1 Introduction

Short implants are increasingly being discussed as a treatment alternative in situations characterized by limited vertical bone height [5].

Compared to the use of standard implants due to biomechanical considerations (e.g., crown-to-implant ratio, C/R) with short implants may result in unfavourable loading conditions and complications, including excessive bone loss and implant failure [20]. Improvements in implant design and surface, together with the use of modified implant insertion techniques, are all aimed at minimizing these risks [15].

### 3.2 Definition of short implants

Implants are commonly referred to as short if their designed intrabony length measures  $\leq 8$  mm with diameters  $\geq 3.75$  mm. Standard implants are those with a length  $> 8$  mm with diameters  $\geq 3.75$  mm [47, 52]. Ultra-short implants are those with lengths  $< 6$  mm [16].

### 3.3 Indications for short implants

Short implants are primarily used to avoid bone augmentation procedures in the maxillary and mandibular posterior segments of partially edentulous patients. They are used when vertical bone volume is limited by anatomical structures (maxillary sinus, mandibular canal), but there is sufficient alveolar ridge width to allow successful use of implant diameters  $\geq 3.75$  mm. They are also used for support of removable overdentures as single or multiple tooth replacements in the anterior jaw [25, 52].

### 3.4 Current observations

Various meta-analyses indicate that there is no difference between the use of short implants in comparison to standard implants with grafting procedures for the marginal bone level development or success rates [8, 9, 24, 25, 37, 39, 56, 62, 66].

Whether there is an advantage to splinting the implants remains unclear [1, 36] [54].

In a limited number of studies, immediate loading has been performed [26, 33, 65]. For immediate loading short implants may be used, but care must be taken to follow specific treatment concepts.

However, the literature shows, that short implants with reduced diameter have a failure rate of up to 10% after 3–5 years [13].

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org



### 3.5 Preventing complications

Some authors have made recommendations on how to avoid complications that are mainly biomechanical in nature. These recommendations include:

- Machined-surfaced, short implants should not be used [42].
- Short implants should only be used when bone quality is favourable [13].
- Restoration with single crowns [3, 27, 43, 58].
- Single short implants with cantilevers should not be used [57].
- Guiding surfaces for lateral movement should be avoided [11].
- Regular occlusal checks are recommended [59].
- Placement at or below bone level with tapered abutment design [29, 38].
- The implant surgeon and restorative dentist must have appropriate training [58].

## 4 Use of angulated implants

### 4.1 Introduction

Angulated standard implant designs or non-angulated ones placed in off-axis (tilted) positions are becoming routine in splinted reconstructions of edentulous jaws as an alternative treatment option to avoid hard-tissue augmentation procedures, but also to increase primary stability for immediate loading procedures with longer implants [11]. These concepts require 4 implants in the mandible and 4 to 6 implants in the maxilla.

The aim of placing implants in a tilted position is to utilize as much bone as possible, while still avoiding vital adjacent structures (e.g., the mental foramen in the mandible or the maxillary sinus in the maxilla). They also increase the surface area for restorative support (through divergent implant axes) [6]. Restorations can be inserted on these implants using angulated abutments.

Modifications of this concept are also used in partially edentulous patients or with a reduced number of implants. The specific treatment protocol varies and individual recommendations should be followed.

### 4.2 Current observations

Based on 24 included articles, 2,637 patients which were rehabilitated with 2,735 full prostheses (1,464 maxillary, 1,271 mandibular), supported by 5,594 and 5,611 tilted and axial implants, respectively. The cumulative implant survival rate between the observation of 3 up-to 18 years was 93.91% and 99.31% for implants and prostheses, respectively [14].

Despite the positive clinical results, the scientific discussion on the clinical relevance of marginal bone level development around angled implants is still ongoing [12, 14, 41, 48].

### 4.3 Restorative experience

The use of a cantilevered, shortened dental arch with a lack of posterior support has not shown an increased prevalence of oromandibular malfunctions [51].

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org



#### 4.4 Preventing complications

- The use of angulated implants splinted with fixed dental prostheses and subjected to immediate loading should achieve adequate primary stability [41, 49].
- Preoperative 3D computer-assisted diagnosis is recommended for anatomically and prosthetically correct angled implant placement [30].
- The implant surgeon and restorative dentist must have adequate training [60].

### 5 Use of reduced-diameter implants

#### 5.1 Definition

Reduced-diameter implants can be defined as those with intraosseous diameters  $< 3.5$  and  $> 2.7$  mm for placement in sites with reduced alveolar ridge bone width. Implants with a diameter  $< 2.7$  mm are referred to as "mini"-implants [21].

#### 5.2 Current observations

Reduced-diameter implants generally have high survival rates ( $> 90\%$ ), assuming careful patient selection, bone density assessment, clinical approach, and user experience [28, 32, 55]. Reduced-diameter implants can also be used in the posterior region with high success rates [32]. These findings are supported by recent meta-analyses [23, 53, 61].

Despite the limited number of studies available, fixed dental prostheses supported by reduced-diameter implants showed comparable survival and success rates to those supported by standard-diameter implants, with slightly lower marginal bone loss. No firm conclusions could be drawn for partial removable dental prostheses [7].

Mandibular mini-implants supporting an overdenture exhibit excellent short- to medium-term survival rates and improve patients' chewing and speaking ability, quality of life, and satisfaction [22, 31, 35, 40]. Survival of mini-implants supporting maxillary overdentures has been observed to be lower, but patients will benefit in terms of oral health-related quality of life [35, 53, 63].

Mini-implants also show favourable results as supporting implants for removable partial dentures [4, 45, 46].

#### 5.3 Preventing complications

- Mini-implants have an increased risk of implant loss in the maxilla.
- Short mini-implants should be avoided [64].
- The implant surgeon and restorative dentist must have adequate training

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org



Guideline: Update on short, angulated and diameter-reduced implants  
18th European Consensus Conference (EuCC), February 2023  
Page 6 of 10

## 6 Recommendations for short, angulated and/or reduced-diameter implants

The use of short, angulated or reduced-diameter implants in sites with reduced bone volume can be a reliable, faster and less risky therapeutic option in terms of specific treatment parameters, compared with the risks associated with the use of standard-dimension implants in combination with augmentation procedures. The implant surgeon and the restorative dentist must have appropriate training to select the best possible therapy for each patient [44].

Cologne, 6 February 2023

Professor Dr. Dr. Joachim Zöller  
Vice President

Professor Dr. Jörg Neugebauer  
Secretary General/Host

BDIZ EDI  
Lipowskyst. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org



## References

1. Afrashtehfar KI, Katsoulis J, Koka S et al. Single versus splinted short implants at sinus augmented sites: A systematic review and meta-analysis. *J Stomatol Oral Maxillofac Surg* 2021; 122: 303-310.
2. Aghaloo TL, Moy PK. Which hard tissue augmentation techniques are the most successful in furnishing bony support for implant placement? *Int J Oral Maxillofac Implants* 2007; 22 Suppl: 49-70.
3. Al-Ansari A. Short implants supporting single crowns in atrophic jaws. *Evid Based Dent* 2014; 15: 85-86.
4. Al Jaghsi A, Heinemann F, Biffar R et al. Immediate versus delayed loading of strategic mini-implants under existing removable partial dentures: patient satisfaction in a multi-center randomized clinical trial. *Clin Oral Investig* 2021; 25: 255-264.
5. Aloy-Prosper A, Penarrocha-Oltra D, Penarrocha-Diago M et al. The outcome of intraoral onlay block bone grafts on alveolar ridge augmentations: a systematic review. *Med Oral Patol Oral Cir Bucal* 2015; 20: e251-258.
6. Aparicio C, Perales P, Rangert B. Tilted implants as an alternative to maxillary sinus grafting: a clinical, radiologic, and periotest study. *Clin Implant Dent Relat Res* 2001; 3: 39-49.
7. Badaro MM, Herdt B, Bezerra AP et al. Narrow-Diameter Implants for Partial Fixed and Removable Prosthesis: A Systematic Review and Meta-Analysis. *Int J Prosthodont* 2022; 35: 738-751.
8. Bitinas D, Bardijevskyt G. Short implants without bone augmentation vs. long implants with bone augmentation: systematic review and meta-analysis. *Aust Dent J* 2021; 66 Suppl 1: S71-S81.
9. Chaware SH, Thakare V, Chaudhary R et al. The rehabilitation of posterior atrophic maxilla by using the graftless option of short implant versus conventional long implant with sinus graft: A systematic review and meta-analysis of randomized controlled clinical trial. *J Indian Prosthodont Soc* 2021; 21: 28-44.
10. Chiapasco M, Casentini P, Zaniboni M. Bone augmentation procedures in implant dentistry. *Int J Oral Maxillofac Implants* 2009; 24 Suppl: 237-259.
11. Chrcanovic BR, Albrektsson T, Wennerberg A. Tilted versus axially placed dental implants: a meta-analysis. *J Dent* 2015; 43: 149-170.
12. Cortes-Breton Brinkmann J, Garcia-Gil I, Pedregal P et al. Long-Term Clinical Behavior and Complications of Intentionally Tilted Dental Implants Compared with Straight Implants Supporting Fixed Restorations: A Systematic Review and Meta-Analysis. *Biology (Basel)* 2021; 10.
13. das Neves FD, Fones D, Bernardes SR et al. Short implants--an analysis of longitudinal studies. *Int J Oral Maxillofac Implants* 2006; 21: 86-93.
14. Del Fabbro M, Pozzi A, Romeo D et al. Outcomes of Fixed Full-Arch Rehabilitations Supported by Tilted and Axially Placed Implants: A Systematic Review and Meta-Analysis. *Int J Oral Maxillofac Implants* 2022; 37: 1003-1025.
15. Deporter D. Short dental implants: what works and what doesn't? A literature interpretation. *Int J Periodontics Restorative Dent* 2013; 33: 457-464.
16. Deporter D, Ogiso B, Sohn DS et al. Ultrashort sintered porous-surfaced dental implants used to replace posterior teeth. *J Periodontol* 2008; 79: 1280-1286.
17. Esposito M, Felice P, Worthington HV. Interventions for replacing missing teeth: augmentation procedures of the maxillary sinus. *Cochrane Database Syst Rev* 2014; 5: CD008397.
18. Esposito M, Grusovin MG, Kwan S et al. Interventions for replacing missing teeth: bone augmentation techniques for dental implant treatment. *Cochrane Database Syst Rev* 2008; CD003607.
19. Esposito M, Grusovin MG, Worthington HV et al. Interventions for replacing missing teeth: bone augmentation techniques for dental implant treatment. *Cochrane Database Syst Rev* 2006; CD003607.
20. Garaicoa-Pazmino C, Suarez-Lopez del Amo F, Monje A et al. Influence of crown/implant ratio on marginal bone loss: a systematic review. *J Periodontol* 2014; 85: 1214-1221.
21. Gleiznys A, Skirbutis G, Harb A et al. New approach towards mini dental implants and small-diameter implants: an option for long-term prostheses. *Stomatologija* 2012; 14: 39-45.
22. Goiato MC, Sonogo MV, Pellizzer EP et al. Clinical outcome of removable prostheses supported by mini dental implants. A systematic review. *Acta Odontol Scand* 2018; 76: 628-637.

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org





23. Gonzalez-Valls G, Roca-Millan E, Cespedes-Sanchez JM et al. Narrow Diameter Dental Implants as an Alternative Treatment for Atrophic Alveolar Ridges. Systematic Review and Meta-Analysis. *Materials (Basel)* 2021; 14.
24. Grunau O, Terheyden H. Lateral augmentation of the sinus floor followed by regular implants versus short implants in the vertically deficient posterior maxilla: a systematic review and timewise meta-analysis of randomized studies. *Int J Oral Maxillofac Surg* 2022.
25. Guida L, Bressan E, Cecoro G et al. Short versus Longer Implants in Sites without the Need for Bone Augmentation: A Systematic Review and Meta-Analysis of Randomized Controlled Trials. *Materials (Basel)* 2022; 15.
26. Hadilou M, Ebrahimi P, Karimzadeh B et al. Immediate loading of short implants: A systematic review. *J Adv Periodontol Implant Dent* 2021; 13: 15-21.
27. Hasan I, Bourauel C, Mundt T et al. Biomechanics and load resistance of short dental implants: a review of the literature. *ISRN Dent* 2013; 2013: 424592.
28. Hasan I, Bourauel C, Mundt T et al. Biomechanics and load resistance of small-diameter and mini dental implants: a review of literature. *Biomed Tech (Berl)* 2014; 59: 1-5.
29. Hentschel A, Herrmann J, Glauche I et al. Survival and patient satisfaction of short implants during the first 2 years of function: a retrospective cohort study with 694 implants in 416 patients. *Clin Oral Implants Res* 2016; 27: 591-596.
30. Jacobs R, Salmon B, Codari M et al. Cone beam computed tomography in implant dentistry: recommendations for clinical use. *BMC Oral Health* 2018; 18: 88.
31. Jawad S, Clarke PT. Survival of Mini Dental Implants Used to Retain Mandibular Complete Overdentures: Systematic Review. *Int J Oral Maxillofac Implants* 2019; 34: 343-356.
32. Klein MO, Schiegnitz E, Al-Nawas B. Systematic review on success of narrow-diameter dental implants. *Int J Oral Maxillofac Implants* 2014; 29 Suppl: 43-54.
33. Kulkarni V, Uttamani JR, Asar NV et al. Evidence-Based Clinical Outcomes of Immediate and Early Loading of Short Endosseous Dental Implants: A Meta-analysis. *Int J Oral Maxillofac Implants* 2021; 36: 59-67.
34. Lee SA, Lee CT, Fu MM et al. Systematic review and meta-analysis of randomized controlled trials for the management of limited vertical height in the posterior region: short implants (5 to 8 mm) vs longer implants (> 8 mm) in vertically augmented sites. *Int J Oral Maxillofac Implants* 2014; 29: 1085-1097.
35. Lemos CA, Verri FR, Batista VE et al. Complete overdentures retained by mini implants: A systematic review. *J Dent* 2017; 57: 4-13.
36. Li QL, Yao MF, Cao RY et al. Survival Rates of Splinted and Nonsplinted Prostheses Supported by Short Dental Implants ( $\leq 8.5$  mm): A Systematic Review and Meta-Analysis. *J Prosthodont* 2022; 31: 9-21.
37. Lin ZZ, Jiao YQ, Ye ZY et al. The survival rate of transcresal sinus floor elevation combined with short implants: a systematic review and meta-analysis of observational studies. *Int J Implant Dent* 2021; 7: 41.
38. Lombardo G, Corrocher G, Pighi J et al. The impact of subcrestal placement on short locking-taper implants placed in posterior maxilla and mandible: a retrospective evaluation on hard and soft tissues stability after 2 years of loading. *Minerva Stomatol* 2014; 63: 391-402.
39. Lozano-Carrascal N, Anglada-Bosqued A, Salomo-Coll O et al. Short implants (<8mm) versus longer implants ( $\geq 8$ mm) with lateral sinus floor augmentation in posterior atrophic maxilla: A meta-analysis of RCT;s in humans. *Med Oral Patol Oral Cir Bucal* 2020; 25: e168-e179.
40. Marcello-Machado RM, Faot F, Schuster AJ et al. Mini-implants and narrow diameter implants as mandibular overdenture retainers: A systematic review and meta-analysis of clinical and radiographic outcomes. *J Oral Rehabil* 2018; 45: 161-183.
41. Mehta SP, Sutariya PV, Pathan MR et al. Clinical success between tilted and axial implants in edentulous maxilla: A systematic review and meta-analysis. *J Indian Prosthodont Soc* 2021; 21: 217-228.
42. Menchero-Cantalejo E, Barona-Dorado C, Cantero-Alvarez M et al. Meta-analysis on the survival of short implants. *Med Oral Patol Oral Cir Bucal* 2011; 16: e546-551.
43. Mezzomo LA, Miller R, Triches D et al. Meta-analysis of single crowns supported by short (<10 mm) implants in the posterior region. *J Clin Periodontol* 2014; 41: 191-213.
44. Monteiro DR, Silva EV, Pellizzer EP et al. Posterior partially edentulous jaws, planning a rehabilitation with dental implants. *World J Clin Cases* 2015; 3: 65-76.

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org



Guideline: Update on short, angulated and diameter-reduced implants  
18th European Consensus Conference (EuCC), February 2023  
Page 9 of 10

45. Mundt T, Al Jaghsi A, Schwahn B et al. Immediate versus delayed loading of strategic mini dental implants for the stabilization of partial removable dental prostheses: a patient cluster randomized, parallel-group 3-year trial. *BMC Oral Health* 2016; 17: 30.
46. Mundt T, Schwahn C, Heinemann F et al. Stabilizing Removable Partial Dentures by Immediate or Delayed Loading of Mini-implants: Chewing Efficiency in a Randomized Controlled Clinical Trial. *Int J Oral Maxillofac Implants* 2020; 35: 178-186.
47. Olate S, Lyrio MC, de Moraes M et al. Influence of diameter and length of implant on early dental implant failure. *J Oral Maxillofac Surg* 2010; 68: 414-419.
48. Omori Y, Lang NP, Botticelli D et al. Biological and mechanical complications of angulated abutments connected to fixed dental prostheses: A systematic review with meta-analysis. *J Oral Rehabil* 2020; 47: 101-111.
49. Penarrocha-Diago M, Penarrocha-Diago M, Zaragoza-Alonso R et al. Consensus statements and clinical recommendations on treatment indications, surgical procedures, prosthetic protocols and complications following All-On-4 standard treatment. 9th Mozo-Grau Ticare Conference in Quintanilla, Spain. *J Clin Exp Dent* 2017; 9: e712-e715.
50. Pommer B, Mailath-Pokorny G, Haas R et al. Patients' preferences towards minimally invasive treatment alternatives for implant rehabilitation of edentulous jaws. *Eur J Oral Implantol* 2014; 7 Suppl 2: S91-109.
51. Reissmann DR, Heydecke G, Schierz O et al. The randomized shortened dental arch study: temporomandibular disorder pain. *Clin Oral Investig* 2014; 18: 2159-2169.
52. Renouard F, Nisand D. Impact of implant length and diameter on survival rates. *Clin Oral Implants Res* 2006; 17 Suppl 2: 35-51.
53. Schiegnitz E, Al-Nawas B. Narrow-diameter implants: A systematic review and meta-analysis. *Clin Oral Implants Res* 2018; 29 Suppl 16: 21-40.
54. Shah AH, Patel P, Trivedi A et al. A comparison of marginal bone loss, survival rate, and prosthetic complications in implant-supported splinted and nonsplinted restorations: A systematic review and meta-analysis. *J Indian Prosthodont Soc* 2022; 22: 111-121.
55. Sohrabi K, Mushantat A, Esfandiari S et al. How successful are small-diameter implants? A literature review. *Clin Oral Implants Res* 2012; 23: 515-525.
56. Terheyden H, Meijer GJ, Raghoobar GM. Vertical bone augmentation and regular implants versus short implants in the vertically deficient posterior mandible: a systematic review and meta-analysis of randomized studies. *Int J Oral Maxillofac Surg* 2021; 50: 1249-1258.
57. Thoma DS, Wolleb K, Schellenberg R et al. Two short implants versus one short implant with a cantilever: 5-Year results of a randomized clinical trial. *J Clin Periodontol* 2021; 48: 1480-1490.
58. Thoma DS, Zeltner M, Husler J et al. EAO Supplement Working Group 4 - EAO CC 2015 Short implants versus sinus lifting with longer implants to restore the posterior maxilla: a systematic review. *Clin Oral Implants Res* 2015; 26 Suppl 11: 154-169.
59. Todescan S, Lavigne S, Kelekis-Cholakis A. Guidance for the maintenance care of dental implants: clinical review. *J Can Dent Assoc* 2012; 78: c107.
60. Unsal GS, Turkyilmaz I, Lakhia S. Advantages and limitations of implant surgery with CAD/CAM surgical guides: A literature review. *J Clin Exp Dent* 2020; 12: e409-e417.
61. Valente NA, Marchio V, Troiano G et al. Narrow-diameter versus standard-diameter implants placed in horizontally regenerated bone in the rehabilitation of partially and completely edentulous patients: A systematic review. *Int J Oral Implantol (Berl)* 2022; 15: 11-33.
62. Vetromilla BM, Mazzetti T, Pereira-Cenci T. Short versus standard implants associated with sinus floor elevation: An umbrella review of meta-analyses of multiple outcomes. *J Prosthet Dent* 2021; 126: 503-511.
63. Vi S, Pham D, Du YYM et al. Mini-Implant-Retained Overdentures for the Rehabilitation of Completely Edentulous Maxillae: A Systematic Review and Meta-Analysis. *Int J Environ Res Public Health* 2021; 18.
64. Wang HL, Okayasu K, Fu JH et al. The success rate of narrow body implants used for supporting immediate provisional restorations: a pilot feasibility study. *Implant Dent* 2012; 21: 467-473.
65. Wu H, Shi Q, Huang Y et al. Failure Risk of Short Dental Implants Under Immediate Loading: A Meta-Analysis. *J Prosthodont* 2021; 30: 569-580.
66. Yu X, Xu R, Zhang Z et al. A meta-analysis indicating extra-short implants ( $\leq 6$  mm) as an alternative to longer implants ( $\geq 8$  mm) with bone augmentation. *Sci Rep* 2021; 11: 8152.

BDIZ EDI  
Lipowskystr. 12  
D-81373 Munich  
GERMANY

Fon: +49-89-720 69 888  
Fax: +49-89-720 69 889  
office@bdizedi.org  
www.bdizedi.org