

Optimized implant planning: DICOM-STL matching

Author_ Dr Frank Schaefer, Dr Dagmar Schaefer & Dr Mike Zäuner, Germany

Introduction

On the basis of three-dimensional X-ray images, in the 1990s the first software programmes allowed a navigated insertion of dental implants. But the digitisation of dental processes started even earlier, namely in the mid-1980's. Imaging techniques allowed the production of components based on virtual construction. Today, this principle is well established both in the dental-clinical field and especially in the dental laboratory. Meanwhile, 3-D data sets of objects are created not only by normal camera shots, but there are also special 3-D scanners in use. In particular, today's desktop scanners are so precise in their resolution accuracy that they are able to exactly reflect the real model or oral situation. Simultaneously with the capturing process, different methods have been developed to transfer the acquired 3-D data sets back to reality. While initially this was a milling and prototyping process, currently

the sintering and printing processes are favoured. For a long time, navigated implantology and 3-D scanning has been developed in parallel, where at best surgical templates were fabricated by prototyping on basis of X-ray data sets.

Goal: optimal implant position

In recent years, the matching of 3-D X-ray data sets (DICOM) and 3-D model data sets (STL) has begun. The goal was and still is to find the optimal surgical and prosthetic implant positions for navigated insertion to provide an optimal solution for the patient. In addition, the production of temporary dentures and in individual cases an immediate treatment is so much better and much more reliable and predictable. At the same time, an objective quality control of both the planning and the result is practicable through matching of DICOM and STL data sets. By means of some case studies, we show which diagnostic and technical possibilities have been feasible since the establishment of the diagnostics and navigation system CTV in 2005 in the following article.


Implant planning with CTV

X-rays are subject to the laws of physics. Therefore, all the resulting images are generally afflicted with an error regarding distortion, diffraction and interference. Because these errors have their origin in the radiological density changes of the object, some areas cannot be represented or are misrepresented. Particularly critical are movement-induced distortions in CBCT images. They cannot be completely avoided or even predicted. A further increase in accuracy solely from radiological data does not seem to be possible currently. The solution is to collect additional data by using independent methods to achieve a "rectification" and detail enhancement through combination with the radiological data. For example, the line of the gingiva and other surface structures in the 3-D X-ray image cannot be traced precisely. The solution here is the correct matching

Fig. 1 _Planned implant with full-guided drill sleeve; gingiva line set by the clinician according to X-ray image (pink); by matching obtained with the situation model – real – gingiva line (yellow).

Fig. 2a _Virtual matching in 3-D X-ray; planned implants with a situation model (red). Image cut: orthogonal ridge section with gingiva line (yellow) from the situation model.

Fig. 2b _Virtual matching in 3-D X-ray, implants planned with aesthetics wax up (green); image: orthogonal ridge section with tooth line (yellow) of aesthetic wax up.



THE PATENTED, REVOLUTIONARY BONE GRAFT COMPOSITE OF THE 21ST CENTURY

Hypro-Oss®

is a revolutionary, natural, atelo-peptidized and lyophilized bovine bone graft composite.

natural bovine bone composite with incorporated atelocollagen in each granule

30% collagen Type I free of telopeptide (atelocollagen) with native osteoinductive elements (TGF, BMPs, IGFs)

70% native osteoconductive bovine Hydroxyapatite components

powerful haemostatic and bacteriostatic effects thanks to native atelocollagen components



excellent handling due to effective hydrophilic properties and optimal cell adhesion

natural crystalline structure and optimal porosity guarantees


long term dimensional stability

new bone formation within the shortest period of time ever

user reported: no swelling or hematoma complications after sinus lifting or other implantation procedures



Histology after 4 weeks.



Histology after 14 weeks.




Complete your portfolio with
our atelo-collagen membranes
and haemostatics



Fig. 3 Post-surgical matching; inserted implant (left), transition with virtual implant (middle), with virtual implant with abutment (right); from gingiva line situation model (yellow).

Fig. 4a STL mesh of situation model with inserted drilling sleeves; drilling sleeves separately (top left).

Fig. 4b STL mesh of situation model with laboratory implant analogs; laboratory implant analogs separately (top left).



of DICOM data sets with the digital capture of the associated surface structures, e.g. anatomical model. With the situation model the real surface profile is obtained. If an aesthetic modelling (wax-up) is scanned and matched additionally, the planned position of the implants both in axial direction as well as in mesial/distal orientation can be determined optimally (Figs. 1, 2a & b).

3-D data matching

The comparison of the real positions of the inserted implants in the jaw with the virtual planning

is done by matching the 3-DX-ray planning capturing with the post-op 3-D picture. Here it is irrelevant whether the planning and the post-op 3-D capturing come from the same device type (DVT/CT) or not. This method also allows for a standardised follow-up (Fig. 3).


DICOM and STL data matching

For the manufacturing of surgical templates, for models to produce temporary restorations in navigated implantations and planning of definite dentures (backward-planning) matching data sets from

Fig. 5 Model replica of SLT-data set with: a) surgical drilling sleeves on fabricated guides, b) designed laboratory implant analogs, c) implemented real abutments.

Fig. 6 Calculated OPG from CBCT data and planned implant positions; arrows: planning under still existing dentures.

Fig. 7 Situation 4th quadrant of Figure 6 including planned implant positions and abutments (parallelised). Mesh of the situation model (red) and wax-up (green).




ORAL IMPLANTOLOGY WORLD CONGRESS

PARIS, July 2nd 2014

Maison de la Chimie

28, rue Saint-Dominique - 75007 Paris

wfld@clq-group.com



UNIVERSITÉ
PARIS DESCARTES

www.oiw-Paris2014.com

Fig. 8a_STL mesh of the situation model of Figure 6 with designed drilling sleeve guides, even under existing dentures (regio 36, 44/45).

Fig. 8b_Replication model resulting from Figure 8a with drill sleeve guides.

Fig. 8c_Model replica with attached surgical drilling guides (Steco) in preparation of the production of surgical drilling template.



DICOM and STL data are used. Virtual models can be designed with exactly positioned sleeves for full-guided systems and or with laboratory analogues of the planned implants. This range can be extended, provided that the STL data sets of components to be designed are available, such as implant abutments. The thus created virtual model is transferred by milling, printing, sintering, etc. back to reality and can then be used e.g. in the laboratory for the production of temporary dentures or surgical guides. The more accurate the replications process the better the models (Figs. 4a–b, 5a–c).

Complex planning

For complex planning, even when there is not an optimal bone situation and accompanying surgical services (e.g. sinus lift) are needed, the matching processes of the CTV system support the surgeon. By virtual articulation of the scanned models and matching with the X-ray data, a position and axial direction of the planned implants and their subsequent supra structure in relation to the remaining dentures or natural teeth are determined and other accompanying, necessary surgical procedures can be pre-planned (Figs. 9a–g).

Safe implant-planning

It is also possible to safe implant-planning make with still incorporated metal structures, even if the X-ray image at these locations with radiation artifact areas is insufficiently evaluable. In the described case, the usage of a non-optimal DVT had been assumed, due to extensive metal restorations. Alternatively, the structures would have to be removed. Because of many opportunities in the CTV system, a virtual planning for minimally invasive, navigated implantation is almost unrestricted. (Figs. 6–8c)

Comprehensive matching process

Last but not least, quality controls, such as of the finished surgical drilling template, are carried out with these comprehensive matching processes. In order to achieve this, the template is scanned and matched as best as possible with the planning images for covering. Ideally, there are no deviations. If differences occur, the implantologist must decide whether he can use this template or a new preparation will be necessary. In this way, failures in implantation and subsequent prosthetic treatment are avoided (Figs. 10a–e, 11a–c).

Fig. 9a_Matching: situation model (red), wax-up (green), the opposing jaw (yellow) and DICOM dataset using virtual articulator; planned implant positions (orange line: gingiva).

Fig. 9b_Hidden situation model; situation as seen in Figure 9a.


Fig. 9c_Situation of Figure 9a: orthogonal slice with planned implant to planned full-guided drilling sleeve; gingiva line (orange); wax-up contour (green).

Fig. 9d_Situation model (red) with the opposing jaw (yellow) in the virtual articulator and the planned implant positions and parallelised abutments.

Fig. 9e_Situation in Fig. 9d, posterior view.

Fig. 9f_Situation model (red) with the opposing jaw (yellow) in the virtual articulator with the planned implant positions, parallelised abutments and position for drilling sleeves (full-guided).

Fig. 9g_STL mesh with drilling sleeves for the production of the replica model (surgical template).



44TH

INTERNATIONAL ANNUAL CONGRESS OF THE DGZI

Been there already?
Concepts in Implantology

September 26–27, 2014
Düsseldorf, Germany | Hilton Hotel

Scientific Director: Dr. Roland Hille/DE

SPEAKERS

Prof. Dr. Florian Beuer/DE
Priv.-Doz. Dr. Kai-Hendrik Bormann/DE
Prof. Dr. Suheil Boutros/US
Prof. Dr. Herbert Deppe/DE
Dr. Dirk U. Duddeck/DE
Prof. Dr. Wolf-D. Grimm/DE
Priv.-Doz. Dr. Friedhelm Heinemann/DE
Prof. Dr. Kai-Olaf Henkel/DE
Prof. Dr. Guido Heydecke/DE
Dr. Detlef Hildebrand/DE
Priv.-Doz. Dr. Dr. Marcus O. Klein/DE
Prof. Dr. Johannes Kleinheinz/DE
Prof. Dr. Regina Mericske/CH
Dr. Dr. Manfred Nilius/DE
Prof. Dr. Dipl.-Ing. Jürgen Richter/DE
Dr. Achim W. Schmidt, M.Sc./DE
Prof. Dr. Dr. Ralf Smeets/DE
Prof. Dr. Thomas Weischer/DE
Prof. Dr. Dr. Richard Werkmeister/DE

IMPRESSIONS
43rd INTERNATIONAL
ANNUAL CONGRESS OF THE DGZI



Gold Sponsor

otmedical®

Silver Sponsor

straumann
simply doing more

Bronze Sponsor

Implant Direct
simply smarter.

DGZI
Deutsche Gesellschaft für
Zahnärztliche Implantologie e.V.

FAX REPLY // +49 341 48474-290


Please send me further information on the
44TH INTERNATIONAL ANNUAL CONGRESS OF THE DGZI
on 26–27 September, 2014, in Düsseldorf, Germany.

Office Stamp

Fig. 10a_STL mesh of the real drilling template (green) and DICOM data with the planned implants maxillary anterior region.

Fig. 10b_Planned implant positions of maxillary anterior area with evaluated positions for drill sleeves (full-guided) and surface line of the scans of the real drilling template (green, regio 11).

Fig. 10c-e_Orthogonal cross-sectional images with planned implants (aqua), associated full-guided drilling sleeves (dark green), orange line: surface profile (gingiva) of situation model, green line surface during the scan of the template: line "based" on the virtual sleeve edges.



Conclusion

The procedures for overlay of DICOM and STL data contained in the CTV system allow a comprehensive planning of implant positions regarding surgical, prosthetic and aesthetic aspects. Due to the diversity of options, shortcomings of X-ray or model data sets can be fairly settled. This method eliminates the need of a special transfer device for the implementation of the design positions from the virtual to the real world. Thus, the described approach is independent from the existing dental infrastructure as the data exchange with freely selectable machining centres can be done via internet. The goal is to enable a consistent minimally invasive

surgical-implantological procedure, to reduce failure rates and to meet the often high demand for prosthetics and aesthetics from the patient's perspective.

contact

implants

Dr Frank Schaefer & Dr Dagmar Schaefer


PraxisSoft Dr. D. Schaefer e. K.
Haarbergstraße 21, 99097 Erfurt, Germany

Tel.: +49 361 3468914 (4230713)
d.schaefer@praxissoft.org

Fig. 11a_STL mesh of situation model (red) and wax-up (green) with the planned implant and abutment.

Fig. 11b_STL mesh with designed drilling sleeves of the replica model for the production of the surgical guide.

Fig. 11c_Orthogonal cross section regio 45 with planned implant and sleeve position; orange line: surface profile situation model (gingiva); green line: surface contour wax-up; aqua line: real surface of drilling template.



INTRODUCING



**DIGITAL
DENTISTRY
SHOW**

AT
INTERNATIONAL EXPODENTAL MILAN

16 17 18 OCTOBER 2014

EXHIBITION

LIVE PRODUCT PRESENTATIONS

HANDS-ON WORKSHOPS

PRINTED REFERENCE GUIDE

COFFEE WITH THE EXPERTS

www.DigitalDentistryShow.com

Organized by Dental Tribune International in cooperation with Promunidi.
Dental Tribune International | Holbeinstraße 29 | 04229 Leipzig | Germany
T +49 341 48474 134 | F +49 341 48474 173
E info@digitaldentistryshow.com | W www.DigitalDentistryShow.com

dti Dental
Tribune
International

PROMUNIDI